

MEASURING EQUIPMENT
FOR INDUSTRIAL WATER



INDEXES	CODE	DESCRIPTION	CONTROL
Hardness	IN-0402-02	Hardness kit for water measurement, by the EDTA compleximetry method, with a range of 0-100°F and a number of reagents 2+1.	Basic Advanced
Hardness	IN-0402-03	REAGENT A	
Hardness	IN-0402-04	REAGENT B	
Hardness	HI-3890	Total Hardness Test Kit: - Method: titration - Range: 0 - 80 TH - Smallest increment: 1 TH	
Chlorine	HI-701	Free Chlorine Mini Photometer with: - Range 0-2.5 ppm - Resolution of 0.01ppm - Reagent HI 93701-01 (powder) / HI 93701 F (liquid)	Essential
TDS THIS Temperat. pH	HI-98129	Waterproof pocket instrument with a range: - TDS from 0 to 2000 ppm - Conductivity from 0 to 3999 µS/cm - Temperature range from 0-60°C/32-140°F - pH from 0.00 to 14.00	Essential
TDS THIS Temperat. pH	HI-98130	Waterproof Pocket Tester: - TDS from 0 to 10 g/L - Conductivity from 0 to 20mS/cm2 - Temperature range from 0-60°C/32-140°F - pH from 0.00 to 14.00	Essential
TDS THIS Temperat. pH	HI-9814	Multiparameter tester with probe: - TDS 0 a 3000ppm - EC 0.00 to 6.00 mS/cm - Temperature from 0 to 60°C - pH 0.00 to 14.00	Essential
pH Temperat.	HI-991001-30	Portable meter with a range of: - pH 2.000 a 20.00 +-2000mV - Temperature -20.0 to 120.0°C/-4.0 to 248.0°F	Advanced
TDS THIS Temperat. Resistivity Salinity	HI-98192	Waterproof measuring instrument with rank: - TDS from 0 to 400 g/L - Conductivity from 0 to 1000 mS/cm - Temperature from -20 to 120°C/-4 to 248°F - Resistivity 0 to 100 MΩ/cm - Salinity 0.0 to 400.0% NaCl	Advanced
pH TDS THIS Temperat. Resistivity Salinity	HI-98195	Waterproof measuring instrument with rank: - pH de 0.00 a 14.00 / +-600mV - ORP +-2000mV - Conductivity from 0 to 400mS/cm - TDS from 0 to 400000 ppm - Temperature -5 to 55°C/23 to 131°F - Resistivity 1.0 to 1000.00kΩcm - Salinity 0.0 to 70 PSU	Advanced

MEASURING EQUIPMENT
FOR INDUSTRIAL WATER



INDEXES	CODE	DESCRIPTION	CONTROL
pH TDS/Cond. Temperat. Resistivity Salinity Oxygen Atm. Pressure Turbidity	HI-98594	Waterproof measuring instrument with a range of: - pH de 0.00 a 14.00 / +-600mV - ORP +-2000mV - Conductivity from 0 to 400mS/cm - TDS from 0 to 400000 ppm - Temp. de -5 a 55°C/23 a 131°F - Resistivity 1.0 to 1000.00kΩcm - Salinity 0.0 to 70 PSU - Dissolved oxygen 0.00 to 500.00 ppm - Pressure atm. 600 to 1133.2 mBar - Turbidity - Oxygen	Advanced
Free Chlorine Total Chlorine pH Isoc Acid.	HI-971044C	Briefcase that i Includes Chlorine, pH, and isocyanuric acid photometers, as well as reagents, calibration standards, accessories, and a carrying case. - Chlorine Free 0.00 to 5.00mg/L - Total chlorine 0.00 to 5.00mg/L - pH 6.5 to 8.0 pH - Isocyanuric acid 0 to 80 mg/L	Basic Advanced
Free Chlorine Total Chlorine pH Isoc Acid. Bromo Iron Iodine	HI-97101C	Briefcase that i Includes Chlorine, pH, and isocyanuric acid photometers, as well as reagents, calibration standards, accessories, and a carrying case. - Chlorine Free 0.00 to 5.00mg/L - Total chlorine 0.00 to 5.00mg/L - pH 6.5 to 8.0 pH - Isocyanuric acid 0 to 80 mg/L - Bromo 0.00 a 10.00mg/L - Iron 0.00 to 1.60 mg/L - Yodo 0.00 a 12.50mg/L	Advanced
Free Chlorine Total Chlorine pH Isoc Acid. Turbidity	HS-94104C	Portable multiparameter turbidimeter and photometer: - Turbidity 0.00-9.99NTU/0.00-50.0 NTU - Chlorine Free 0.00 to 5.00mg/L - Total chlorine 0.00 to 5.00mg/L - pH 6.5 to 8.5 pH - Isocyanuric acid 0 to 80 mg/L	Advanced
Turbidity	HS-93703POOL	ISO 7027 portable turbidimeter with data memory and computer connection: - Turbidity 0.00-50.0NTU/50-1000 FTU	Advanced
pH Absorbance Alkalinity Bromo Chlorine Copper Isoc Acid. Hardness Iron Ozone Nitrate Phosphate	HI-83326-02	Langelier Indicator: - pH 6.5 to 8.5 pH; pH electrode -2.00 to 16.00 pH - Absorbance: 0.000 to 4.000 Abs - Alkalinity 0 to 500 mg/L (CaCO ₃) - Bromo 0.00 a 10.00 mg/L (Br ₂) - Free chlorine 0.00-5.00 mg/L; Total chlorine 0.00-5.00 mg/L - Total copper 0.00 to 5.00 mg/L - Isocyanuric acid 0 to 80 mg/L - Calcium hardness 0 to 2.70 mg/L - Iron 0 to 5.00 mg/L - Ozone 0 to 2.00 mg/L - Nitrate 0 to 30.0 mg/L - Phosphate 0 to 2.50 mg/L	Advanced
Turbidity Cl pH Conduc-tive. Temperat.	HS-98713CLPH-CET	Briefcase for measuring drinking water Parameters required by RD 3/2023. Includes : Portable Turbidimeter - HI98713 // Free Chlorine Pocket Meter - Checker HI701 / . Water Color Pocket Meter - Checker HI727 // pH Pocket Meter - Tester HI98107 // Pocket Conductivity Meter - Tester HI98304 // Compact thermometer with penetration probe - Checktemp HI98501 .	